



Review

Advancing dental implant surface technology – From micron- to nanotopography

Gustavo Mendonça^{a,b}, Daniela B.S. Mendonça^{a,b}, Francisco J.L. Aragão^{a,c}, Lyndon F. Cooper^{b,*}

^a Universidade Católica de Brasília, Pós-Graduação em Ciências Genômicas e Biotecnologia, SGAN Quadra 916,

Av. W5 Norte 70.790-160 Brasília, DF, Brazil

^b Bone Biology and Implant Therapy Laboratory, Department of Prosthodontics, University of North Carolina at Chapel Hill, 404 Brauer Hall, CB #7450, Chapel Hill, NC 27511, USA

^c Embrapa Recursos Genéticos e Biotecnologia, Laboratório de Introdução e Expressão de Genes, PqEB W5 Norte, 70770-900 Brasília, DF, Brazil

ARTICLE INFO

Article history:

Received 14 February 2008

Accepted 11 May 2008

Available online 9 July 2008

Keywords:

Nanotopography
Dental implant
Surface treatment
Sol–gel techniques
Bone regeneration
Cell signaling

ABSTRACT

Current trends in clinical dental implant therapy include use of endosseous dental implant surfaces embellished with nanoscale topographies. The goal of this review is to consider the role of nanoscale topographic modification of titanium substrates for the purpose of improving osseointegration. Nanotechnology offers engineers and biologists new ways of interacting with relevant biological processes. Moreover, nanotechnology has provided means of understanding and achieving cell specific functions. The various techniques that can impart nanoscale topographic features to titanium endosseous implants are described. Existing data supporting the role of nanotopography suggest that critical steps in osseointegration can be modulated by nanoscale modification of the implant surface. Important distinctions between nanoscale and micron-scale modification of the implant surface are presently considered. The advantages and disadvantages of nanoscale modification of the dental implant surface are discussed. Finally, available data concerning the current dental implant surfaces that utilize nanotopography in clinical dentistry are described. Nanoscale modification of titanium endosseous implant surfaces can alter cellular and tissue responses that may benefit osseointegration and dental implant therapy.

© 2008 Elsevier Ltd. All rights reserved.

1. Introduction

Current dental implant success has evolved from modest results of the middle of the past century. Beginning in the late 1960s the focused efforts of P.I. Branemark led to the detailed microscopic characterization of interfacial bone formation at machined titanium endosseous implants [1,2]. These concepts of osseointegration focused the profession on a proscribed surgical technique and the biocompatible nature of the machined titanium surface. Bone formation at the endosseous implant surface was considered a positive outcome that was contrasted to fibrous encapsulation, a negative and undesired result [3]. The main clinical advantage of osseointegration was the predictable clinical result that occurred when an osseous interface was reproducibly formed and maintained at the titanium surface of load bearing dental implants [4].

Over two decades later, osseointegration is widely accepted in clinical dentistry as the basis for dental implant success. The low rate of implant failure in dense bone of the parasymphysal

mandible [5–8] has not been fully recapitulated by subsequent data from studies involving more challenging clinical situations [9,10]. Anecdotal reports of difficulty in achieving high rates of implant success in selected patient populations (e.g. smokers and diabetics) were supported by initial reports [11–13]. The cause of these failures, while not precisely determined, was largely attributed to a failure in bone formation in support of osseointegration. Challenging osseointegration with new protocols such as immediate placement and immediate loading may require further control of bone formation and osseointegration [9].

Failure to achieve osseointegration at a high rate can be attributed to one or more implant, local anatomic, local biologic, systemic or functional factors [5,8]. Clinical control of all of these factors is represented by multidisciplinary treatment planning procedures. While it is presently acknowledged that these, as well as clinician-related factors, are important determinants of endosseous implants success, a major interest in implant design factors is evident and clinical efforts to improve implant success have been focused on increasing the amount of bone that forms at the endosseous implant surface.

Implant surface character is one implant design factor affecting the rate and extent of osseointegration [14–18]. The process of osseointegration is now well described both histologically and at

* Corresponding author. Tel.: +1 919 966 4579; fax: +1 919 966 3821.

E-mail addresses: gmendonca@ufu.br (G. Mendonça), lyndon_cooper@dentistry.unc.edu (L.F. Cooper).

the cellular level. The adhesion of a fibrin blood clot and the population of the implant surface by blood-derived cells and mesenchymal stem cells is orchestrated in a manner that results in osteoid formation and its subsequent mineralization [19–21]. A seamless progression of changing cell populations and elaboration and modification of the tissue/implant interface eventually results in bone forming in direct contact with the implant surface. Precisely how much of the implant surface directly contacts bone, how rapidly this bone accrual occurs, and the mechanical nature of the bone/implant connection is influenced by the nature of the implant surface itself [22].

The character of the implant surface is implicated in this complex process of osseointegration in a number of different ways. Early investigations revealed the biocompatible nature of the cpTitanium implant [23], and revealed some pragmatic advantages for cpTitanium over other suitable materials [24]. Molecular investigations have contributed to defining cellular responses to titanium as “compatible” and advantageous. For example, Suska and colleagues [25] showed relatively low inflammatory signaling within cells in tissues adjacent to cpTitanium implants and suggested that this is a part of the osseointegration process. During the first 10–20 years of applied endosseous implant experience, the concept that cpTitanium implant biocompatibility supported clinical osseointegration success dominated clinical thinking. Subsequently, experiments with surface topography encouraged new considerations of improvements in bone formation at the implant surface.

2. Micron-scale surface topography

The significance of micron-scale topography was highlighted in an important report by Buser and colleagues [26] that compared various surface preparations of cpTitanium to an electropolished surface negative control and a hydroxyapatite coated positive control group. The observation that a micron-scale rough surface prepared by grit blasting and subsequent acid etching was capable of rapid and increased bone accrual reiterated an earlier report that a TiO₂ grit blasted surface also supported more rapid and increased bone accrual at cpTitanium implants [27]. These early observations indicated that the cpTitanium surface could be modified to enhance bone accrual and suggested that cpTitanium was not only “bio-inert” or “biocompatible”, but could influence cellular activity or tissue responses leading to greater osteogenesis.

At least three different lines of thinking have evolved to better interpret or explain how surface topography at the micron-scale can increase bone-to-implant contact. One is the biomechanical theory of Hansson and Norton [28], the second is the concept of contact osteogenesis [29], and the third is a surface signaling hypothesis supported by many cell culture investigations [14,30,31].

Hansson has elegantly described the theoretical interaction of bone with the implant surface and mathematically defined the role of surface roughness at the micron-scale within this hypothetical construct [28]. The result of the theoretical calculations – that an implant surface should be densely covered with pits of approximately 1.5 µm depth and 3–5 µm diameter – is supported by data collected in a series of studies on implant topography effects on bone-to-implant contact [32,33]. There is an appreciation that mechanical interlocking of bone is essential to the improved performance of endosseous implants. One possible explanation is given by the adaptation of bone to mechanical loading played by the osteocytes acting as mechanosensors [34,35]. Evidence of the important relevance of increased bone-to-implant contact has been provided by measurement of the physical interaction of micron-level rough implants with bone using push-out or torque removal assays [36,37]. What has not been fully elucidated is how mechanical signaling in the unmineralized tissue of forming bone and

adjacent connective tissue is affected by the implant surface. The bonding of bone to the implant surface is not implicated as a mechanism of enhancing the early physical associations of the implant with bone.

A principal role for fibrin clot stabilization by the implant surface exemplifies one role that microscale surface roughness may play in improved osseointegration [38]. Described is a physical interlocking of fibrin fibers with the surface features which promotes the directed ongrowth of bone forming cells directly at the implant/bone interface. Topographic enhancement may aid in stabilization of fragile extracellular matrix scaffolds for conduction of cells toward and onto the implant surface (contact guidance) [39].

Several investigators have further described surface topography-specific effects on titanium-adherent osteoblastic cell behavior [40–43]. The overriding theme of these investigations is that surface adhesion-mediated control of cell function underscores the positive influences on bone formation. Many investigations have contributed to the understanding that there is a range of micron-level surface topography that enhances the adherent osteoblasts' differentiation and extracellular matrix formation/mineralization [44]. Together these investigations have shown that increased surface topography effectively enhances extracellular matrix synthesis of adherent cells and provides a faster and more reliable osseointegration response [43,45–57].

A clearly defined role for extracellular matrix proteins–receptors (integrin) has been proposed to transduce topography-specific signals to the adherent cells [40]. One possible way that topography may alter cellular differentiation is through imposed changes in cell shape [58]. Micron-level topography effects on increased bone-to-implant contact are observed in vivo [26,59], and in human clinical histology [60]. Limited evidence that integrins are involved in cellular responses to implant surfaces has been obtained using MG63 cell culture studies [61].

Micron-scale topographic modification of the cpTitanium surface is accepted in the endosseous dental implant marketplace [32,33]. The belief that micron-level surface topography results in greater accrual of bone at the implant surface is supported by some clinical evidence [62,63]. Yet, these surfaces have been generally interpreted to be biocompatible devices with limited ability to directly affect the initial fate of surrounding tissues (e.g. impose bone formation or prevent bone resorption).

Today, a growing aspect of endosseous implant surface research is focused on further enhancing the activity of bone forming cells at the tissue implant interface. This desire for “bioactivity” has been addressed using a variety of different approaches. Clearly, cpTitanium surfaces can be modified to direct specific cellular responses such as osteogenesis. More specifically, cpTitanium implant surfaces can be made to direct the osteoinduction of adherent progenitor cells. While one approach is the immobilization of bioactive peptides or growth factors and notably the BMPs [64,65], other approaches have embraced the use of nanoscale surface engineering to induce intrinsic osteoinductive signaling of the surface adherent cells. The purpose of this review is to explore how nanotechnology applications to the cpTitanium implant surface may provide new opportunities to create endosseous implant surfaces with greater specific control of adherent cell and adjacent tissue fate.

3. Nanotechnology and surface science

Nanotechnology has been defined as “the creation of functional materials, devices and systems through control of matter on the nanometer length scale (1–100 nm), and exploitation of novel phenomena and properties (physical, chemical, and biological) at that length scale” (National Aeronautics and Space Administration). Nanotechnology involves materials that have a nano-sized

topography or are composed of nano-sized materials. These materials have a size range between 1 and 100 nm (10^{-9} m) (Fig. 1). Nanotechnology often involves one-dimensional concepts (nanodots and nanowires) or the self-assembly of more complex structures (nanotubes). Materials are also classified according to their form and structure as nanostructures, nanocrystals, nanocoatings, nanoparticles, and nanofibers [66].

Application of nanotechnology to the dental implant surface involves a two dimensional association of surface features (across and away from the mean surface plane) (Fig. 2). These nanofeatures can be arranged in an organized manner (isotropic) or unorganized manner (anisotropic), often depending on the method of manufacture. Of the surface topographies that have been applied to a dental implant surface, the topography is often characteristically anisotropic. Isotropic features such as nanogrooves or nanopits that are created largely by optical methods are not readily applied to complex screw shaped objects. When these concepts are applied to the endosseous implant surface, implied is the embellishment of the surface with nanometer-scale features that lead to novel physicochemical behavior (e.g. bone bonding) or biochemical events (e.g. altered protein adsorption, cell adhesion with changes in cell behavior).

Nanoscale modification of the titanium endosseous implant surface may affect both the topography as well as the chemistry of the surface. Specific chemical modification of cpTitanium could be

the targeted goal of nanoscale modification. In fact, a complicating feature of nanoscale manipulation of any material is that there are inherent chemical changes of the bulk material surface. Albrektsson and Wennerberg [32] divided implant surface quality into three categories: (1) mechanical properties, (2) topographic properties, and (3) physicochemical properties. They indicated that these characteristics are related and by changing any of these groups the others will also be affected. This important observation is likely to be even more relevant to the discussions of nanotopographic modifications of the endosseous cpTitanium surface. One frequently encountered limitation to studies comparing nano- and micron-level surface topography is that it can be extremely difficult to isolate chemistry or charge effects induced by the nanotopography. When atomic level control of material assembly is approached, the surface properties are influenced by quantum phenomena that do not govern traditional bulk material behavior [67]. It is very difficult but important to distinguish distinct topography-specific effects from allied changes in surface energy or chemical reactivity.

Nanotechnology requires novel ways of manipulating matter in the atomic scale. Several approaches are currently prevalent in the experimental application to endosseous implants (Table 1). One approach involves the physical method of compaction of nanoparticles of TiO_2 vs micron-level particles to yield surfaces with nanoscale grain boundaries [54]. An advantage of this method is




Fig. 1. Nanoscale in perspective. The scanning electron micrograph at 5000 \times (A) fails to represent true nanoscale features of a titanium implant surface. 100,000 \times image (D) shows the complex nanoscale surface; here produced by titania sol-gel deposition.




Fig. 2. Nanoscale surface modification. (A) Self-assembled monolayers (SAMs) can change the topography and chemistry of a surface to impart novel physical and/or biochemical properties. (B) Deposition or chemical modification techniques can apply nanoscale features ($x \leq 100$ nm) in a manner that are distributed in micron-scale ($y > 100$ nm). (C) Other deposition or compaction methods can place nanoscale features in nanoscale distribution. The cell response to surfaces represented by (B) or (C) may be different. (D) Isotropic surfaces can be created in the nanoscale ($x \leq 100$ nm) by subtractive or additive methods. The distribution can be in either the nano- (y) or micron-scale (y'). It is thought that some nanosurfaces mimic natural cell environments.

Table 1
Methods for creating nanofeatures on cpTitanium implants

Methods	Characteristics
<i>Self-assembly of monolayers</i>	The exposed functional end group could be a molecule with different functions (an osteoinductive or cell adhesive molecule).
<i>Physical approaches</i>	
Compaction of nanoparticles	Conserves the chemistry of the surface among different topographies. Not readily applied over implant surfaces.
Ion beam deposition	Can impart nanofeatures to the surface based on the material used.
<i>Chemical methods</i>	
Acid etching	Combined with other methods (sandblasting and/or peroxidation) can impart nanofeatures to the surface and remove contaminants.
Peroxidation	Produces a titania gel layer. Both chemical and topography changes are imparted.
Alkali treatment (NaOH)	Produces a sodium titanate gel layer allowing hydroxyapatite deposition. Both chemical and topography changes are imparted.
Anodization	Can impart nanofeatures to the surface creating a new oxide layer (based on the material used).
<i>Nanoparticle deposition</i>	
Sol–gel (colloidal particle adsorption)	Creates a thin-film of controlled chemical characteristics. Atomic-scale interactions display strong physical interactions.
Discrete crystalline deposition	Superimposes a nanoscale surface topographical complexity on the surface.
<i>Lithography and contact printing technique</i>	Many different shapes and materials can be applied over the surface. Approaches are labor intensive and require considerable development prior to clinical translation and application on implant surface.

that it conserves the chemistry of the surface among different topographies.

Second is the process of molecular self-assembly. Self-assembled monolayers (SAMs) are formed by the spontaneous chemisorption and vertical close-packed positioning of molecules onto some specific substrata, exposing only the end-chain group(s) at the interface [68]. The exposed functional end group could be an osteoinductive or cell adhesive molecule. An example of this is the use of cell adhesive peptide domains (RGD domains) appended to SAMs composed of polyethylene glycol (PEG) and applied to the titanium implant surfaces [69].

A third method is the chemical treatment of different surfaces to expose reactive groups on the material surface and create nanoscale topography. This is popular among current dental implant investigators. NaOH treatment catalyzes the production of titanium nanostructures outward from the titanium surface [70]. Treatment with a NaOH solution produces a sodium titanate gel layer on the Ti surface while H₂O₂ produces a titania gel layer. The NaOH treatment creates a gel-like layer over the material allowing hydroxyapatite deposition. This behavior has also been seen with other metals such as zirconium and aluminum [71–73]. Titanium oxide nanotubes chemically treated with NaOH accelerated HA crystal growth in a simulated body fluid (SBF) [48]. The kinetics of HA formation is significantly accelerated by the presence of the nanostructure associated to the NaOH treatment. Both chemical and topography changes are imparted.

Chemical treatments (peroxidation (H₂O₂) or acid oxidation, such as hydrofluoric acid) have also been used to create nanotopography [15,72,73]. The use of H₂O₂ with acid etching has been shown to create novel nanostructures of amorphous titanium oxide on the implant surface [74]. It was found that the treatment of the implant surface with H₂O₂/HCl increased the adsorption of RGD peptides onto the surface followed by passivated surfaces (30% HNO₃) and heat-treated surfaces [75]. These surface treatments also increased the mineralization in the same order. Treatment with hydrofluoric acid also creates discrete nanostructures on TiO₂

grit blasted surfaces [76]. Several cell culture studies [41,77,78], preclinical investigations [46,79], and clinical studies [18] support the observation that hydrofluoric acid treatment of TiO₂ grit blasted titanium implants is associated with rapid bone accrual at the implant surface. Complex chemical changes induced by these methods may require careful inspection.

The deposition of nanoparticles onto the titanium surface represents a fourth approach to imparting nanofeatures to a titanium dental implant [80]. Sol–gel transformation techniques achieve deposition of nanometer-scale calcium phosphate accretions to the implant surface [81,82]. Alumina, titania, zirconia and other materials can also be applied [83]. Owing to their resultant atomic-scale interactions, the accretions display strong physical interactions [80,84–86]. In a modified approach, Nishimura and colleagues [87] recently demonstrated a directed approach to assembly of CaPO₄ nanofeatures on dual acid-etched cpTitanium implant surfaces. The deposition of discrete 20–40 nm nanoparticles on an acid-etched titanium surface led to increased mechanical interlocking with bone and the early healing of bone at the endosseous implant surface in a rat model.

One of the main concerns related to coating the implant surface is the risk of coating detachment and toxicity of related debris. This question was addressed by Gutwein and Webster [47] who evaluated the relationship of particle size and cell viability and proliferation compared to micron-particles. Nanoparticles of titania and alumina had less negative impact in cell viability and proliferation. There may be an advantage to nanoscale modification of surfaces using sol–gel coating methods. The quantum interaction of high electron density at the atomic level can enforce high bond strength between the substrate and nanoscale coating. Examples of this have been reported for the calcium phosphate (CaP)/discrete crystalline deposition (DCD) sol–gel coating of Ti alloy implant surfaces [88].

A fifth approach to creating nanoscale topography on Titanium is the use of optical methods (typically lithography) reliant on wavelength specific dimensions to achieve the appropriate

nanoscale modification [70]. These approaches are labor intensive methods that require considerable development prior to clinical translation. The present use of lasers to promote micron-level groove on an implant surface can produce micron-level, not nanoscale, modification of the implant surface [89]. Another method of depositing nanoscale material on to the implant surface involves ion beam deposition (e.g. hydroxyapatite) [90]. All are relevant to the endosseous dental implant surface and experimental examples of each can be identified (below).

Nanotopography has been shown to influence cell adhesion, proliferation, differentiation and cell specific adhesion. Related changes in chemistry and nanostructure impart important chemical changes and permit biomimetic relationships between alloplastic surfaces and tissues. It is speculated that alloplastic nanosurfaces possess topographic elements scaled to naturally occurring substrates.

4. Biomimetics and nanotechnology

The recapitulation of natural cellular environments can be achieved at the nanoscale. Nanoscale modification of an implant surface could contribute to the mimicry of cellular environments to favor the process of rapid bone accrual. For example, cell adhesion to basement membranes is an often cited example of nanoscale biomimetics. The structure of the epithelial basement membrane contains pores approximating 70–100 nm [91]. It is suggested that the surface roughness of bone is approximately 32 nm making it within the nanoscale range of current nanotechnology investigations [92–95]. These in vivo examples further exemplify an anisotropic arrangement of nanofeatures. Intentionally placing molecular structures at such resolution on an endosseous implant may be achieved with anisotropic arrangements. The result may be changes in physical properties including enhanced magnetic, catalytic, optical, electrical, mechanical, and biological properties when compared to conventional formulations of the same material [96].

5. Nanotopography alters cellular responses

Surface nanotopography appears to affect cell interactions at surfaces and alter cell behavior when compared to conventional sized topography (Fig. 3) [97–99]. Different physical relationships exist between cells and nano- vs cell and micron-scale surface features. Nanotopography specific effects on cellular behavior have been demonstrated using a wide range of different cell types




Fig. 3. Depiction of broad range of nanoscale topography effects observed in cellular protein adsorption is altered by nanoscale modification of bulk material. Both cell specificity and extent of cell adhesion are altered. Depending on the nano-architecture cell spreading may be increased or decreased. By presently undefined mechanisms, cell proliferation appears to be enhanced by nanoscale topography. For osteoblast, several investigators have shown nanoscale topography enhances osteoblast differentiation.

including epithelial cells, fibroblasts, myocytes and osteoblasts. Nanostructured surfaces possess unique properties that alter cell adhesion by direct (cell–surface interactions) and indirect (affecting protein–surface interactions) mechanisms. Evidence has been gathered using several models and surface systems (Tables 2 and 3).

5.1. Protein/surface interactions – surface wettability

The changes in initial protein–surface interaction are believed to control osteoblast adhesion [108]. This is a critical aspect of the osseointegration process. When implants come into contact with a biological environment, protein adsorption (e.g. plasma fibronectin) that occurs immediately will mediate subsequent cell attachment and proliferation. Cell binding to protein domains of adhesive extracellular matrix proteins involves receptors termed integrin receptors that transmit signals through a collection of proteins on the cytoplasmic face of the contact, termed focal contacts [130]. Surface effects are often mediated through integrins that bind the RGD motif in cell attachment proteins [131]. The RGD motif of cell adhesive proteins such as fibronectin or vitronectin is important in mediating cell adhesion of osteoblasts and other cells to synthetic material surfaces [132]. Nanofeatures could alter the conformation of these RGD containing proteins, a phenomenon known to affect cell adhesion and behavior [133].

Changing the surface energy or wettability of a biomaterial represents a classical approach to altering cell interactions with the surface. Extracellular matrix protein adsorption onto surfaces (to subsequently modulate cell adhesion) is dramatically affected by surface energy. Interestingly, many studies of self-assembled monolayers (SAMs) have demonstrated that hydrophobic groups are more likely to adsorb albumin and that albumin is not replaced by ECM proteins, blocking cell adhesion. Hydrophobic surfaces adsorbed fibrinogen [134], while hydrophilic surfaces allowed an interchange of adsorbed albumin by ECM proteins [135].

Nanoscale topography is a powerful way of altering protein interactions with a surface. Webster and colleagues [53,109] observed an increased vitronectin adsorption on nanostructured surfaces when compared to conventional surfaces. They also found an increased osteoblast adhesion when compared to other cell types, such as fibroblasts, on the nanosurfaces [109]. Another study suggested higher adsorption of fibronectin on hydrophilic SAMs surfaces with greater focal adhesion formation (integrin binding) evident in the osteoblast cells adhered to the hydrophilic SAM treated surfaces [68]. Lim and colleagues [93] more directly related protein adsorption, cell adhesion and the active process of attachment by measurement of increased focal adhesion kinase (FAK) activity. In a study using SAMs biofunctionalized with RGD, Cavalcanti-Adam and colleagues [133] also found that the spacing among the nanofeatures modulates focal adhesion (FA) formation; cells cultured on a 58 nm nanopattern formed normal FA, whereas those plated on a 108 nm nanopattern failed to develop FA. Surface roughness at the nanoscale is an important determinant of protein interactions that ultimately direct cell activity in control of tissue formation at implant surfaces [136].

5.2. Cell adhesion, spreading and motility

Irrespective of the surface-adsorbed proteins, cells are remarkable in their ability to sense nanostructure (Fig. 4). Nanofeatures of a surface affect both cell adhesion and cell motility. Both of these cell traits are attributed, in part, to the function of integrins. Underlying substratum topography influences cell behaviors by both direct and indirect interactions [137]. Indirect interactions are enacted by the interposed adherent proteins described above. Direct interactions involving the integrin receptors with the surface

Table 2
Reported osteoblast responses to nanosurfaces – in vitro

Size/nanofeature	Cell response	Material/fabrication	Cell culture model	Ref.
14, 29, 45 nm nanopits	Change in signaling	Poly(L-lactic acid) and polystyrene (50/50 w/w)/polymer demixing	hFOB	[94]
Ion beam coating thickness ~60 nm SG coating thickness of 70 nm	Change in signaling	Ti ₆ Al ₄ V/ion beam implantation of Zn or Mg or SG coating with HA	Human bone-derived cells	[100]
12 nm ridges/0.2–2 mm separation	Changes in cell cytoskeleton	Ti/PLD	Osteoblast – rat calvaria	[101]
Pits with 120 nm Ø, spacing of 300 nm in orthogonal or hexagonal arrangement	Changes in cell cytoskeleton	PMMA/EBL in silica	hMSCs	[102]
Pits with 120 nm Ø. The pitch between the pits was 300 nm. Hexagonal and square pit arrangements	Changes in cell cytoskeleton – restriction of spreading – filopodia	PMMA/EBL in poly(carbonate)	hMSCs	[103]
Alumina (23-nm average Ø), titania (32-nm average Ø)	Decreased apoptosis	Particles diluted in growth media at concentrations of 10,000, 1000, and 100 mg/ml as well as 10,000, 5500, and 1000 mg/ml	Human osteoblasts	[47]
RMS roughness values from 0.5 to 13 nm	Decreased proliferation	Gradients of polymer crystallinity were fabricated on films of poly(L-lactic acid)/gradient in annealing temperature.	Osteoblast – MC3T3-E1	[104]
0.5–2.4 µm – Ti 0.5–1.4 µm – Ti ₆ Al ₄ V 0.2–0.4 µm – Co ₂₈ Cr ₆ Mo	Increased adhesion	Ti, Ti ₆ Al ₄ V, and CoCrMo alloys/compaction	Human osteoblasts	[54]
7–40 nm	Increased adhesion	Nobium oxidation of cpTi/sol–gel coating	Osteoblast – MC3T3-E1	[105]
HA, Ti-coated HA annealed in air, and Ti-coated HA annealed in N ₂ + H ₂ possessed Sq of 5, 32, and 28 nm, respectively	Increased adhesion	HA/compaction/Ti coating (CaTiO ₃)	Human osteoblasts	[106]
Nanograined/not shown	Increased adhesion	HA, TCP, or CaTiO ₃ /compaction	Human osteoblasts	[107]
nm HA and HA functionalized with RGD	Increased adhesion	Sintering	Human osteoblasts	[108]
Alumina (23-nm Ø diameter), titania (32-nm Ø diameter)	Increased adhesion	Titania or alumina powders/compaction	Osteoblasts from neonatal rat calvaria	[53]
Alumina (24 and 45 nm average Ø), titania (39 and 97 nm average Ø) and HA (67-nm) powders	Increased adhesion	Titania, alumina or HA powders/compaction	Osteoblasts from neonatal rat calvaria Fibroblasts	[109]
Nanotubes of 3.4 ± 0.3 nm	Increased adhesion	cpTi coated with helical rosette nanotubes featuring lysine side chains (HRN-K1)	Human fetal osteoblast	[110]
Self-assembled nanowires 50–100 wide	Increased adhesion	Ti mesh/NaOH treatment	MSCs and mice	[111]
Alumina nanofibers with 2 nm in Ø and ~50 nm in length alumina nanospherical grain size (<100 nm) powder	Increased adhesion – Ca deposition	Alumina grain or nanofibers/compaction	Human osteoblasts	[49]
5–50 nm pores	Increased adhesion – Ca deposition	Ti ₆ Al ₄ V/H ₂ SO ₄ /H ₂ O ₂ 70/30% followed by coating of TiO ₂	Osteoblast – MC3T3-E1	[112]
Nanophase titania (32-nm average Ø) powders	Increased adhesion – Ca deposition	PLGA mixed with titania (in various proportions)/cured in air	Human osteoblasts	[113]
11–85 nm	Increased adhesion – differentiation	Polystyrene–polybromostyrene/polymer demixing	hFOB	[93]
	Increased adhesion – differentiation – Ca deposition	Nanophase titania/(PLGA) composites	Human osteoblasts	[67]
~ 100 nm/nanotubes	Increased Adhesion – Proliferation – Differentiation	titania/anodization	primary rat bone marrow MSCs	[114]
~ 100 nm/nanopores	Increased adhesion – proliferation – differentiation – Ca deposition	Alumina sheets/anodization	Primary murine bone marrow MSCs	[115]
~ 100 nm features on Ti	Increased differentiation	cpTi/TiO ₂ blasting/HF treatment	Osteoblast – MC3T3-E1 and Ratus novgicus	[78]
100–200 nm difference among groups	Increased differentiation	PMMA/Colloidal lithography and polymer demixing	Primary human osteoprogenitors	[116]

(continued on next page)

Table 2 (continued)

Size/nanofeature	Cell response	Material/fabrication	Cell culture model	Ref.
20–50 nm surface features	Increased differentiation	cpTi and Ti ₆ Al ₄ V/oxidation with H ₂ SO ₄ /H ₂ O ₂	Primary rat calvaria derived osteoblasts	[117]
Elongated HA nanocrystals, with a mean length of about 100 nm	Increased differentiation	Ti ₁₃ Nb ₁₃ Zr/mechanomaking process or Ti ₆ Al ₄ V followed by HF/HNO ₃ acid etch CaP coating	hMSCs	[118]
Parallel ridges/channels (microstructured)/ nanostructured HA (100 nm).	Increased differentiation	Photolithography/nanostructured HAP (biomimetic) on silicon microstructures	Saos-2 and MG63 cell lines	[119]
Alumina nanofibers with 2 nm in \varnothing and ~50 nm in length	Increased differentiation – Ca deposition	Alumina nanofibers/compaction/Sintered at 400, 600, 800, 1000, or 1200 °C	Human osteoblast	[56]
20–50 nm surface features	Increased differentiation – Ca deposition	cpTi/oxidation with H ₂ SO ₄ /H ₂ O ₂	Primary rat calvaria derived osteoblasts	[120]
Alumina (24-nm average \varnothing), titania (39-nm average \varnothing) and HA (67-nm) powders	Increased differentiation – Ca deposition	Titania, alumina or HA powders/compaction	Osteoblasts from neonatal rat calvaria	[52]
Island height of about 90 nm	Increased filopodia	Polystyrene and polybromostyrene/polymer demixing	Human bone marrow cells	[121]
Nanofibers (60–100 nm)	Increased osteoblast specificity	Carbon nanofibers/compaction	Human osteoblasts	[50,122]
Alumina (23-nm average \varnothing), titania (49-nm average \varnothing) and HA (67-nm) powders	Increased osteoblast specificity	Poly(L-lactic) acid or PMMA powder mixed with titania, alumina or HA (in various proportions)/compaction	Neonatal rat calvaria osteoblasts Rat skin fibroblasts	[123]
Nanophase titania (32-nm average \varnothing) powders	Increased osteoblast specificity	PLGA mixed with titania (in various proportions)/cured in air	Human osteoblasts	[124]
~160 nm pores	Increased proliferation	Alumina/EBE	Human osteoblasts	[125]
AAT texture showed micropores and an overlapped nanometric net of filaments	Increased proliferation	cpTi/alkali etching process with CaP solution (biomimetic)	Osteoblast-like MG63	[126]

cpTi – commercially pure titanium, PLGA – poly(lactic-co-glycolic) acid, EBL – electron beam lithography, EBE – electron beam evaporation, HF – hydrofluoric acid treatment, PLD – pulsed laser deposition, PMMA – polymethyl methacrylate, SG – sol-gel, and Ti – titanium.

may also transmit signals to control adhesion, spreading and motility.

Nanofeatures of an alloplastic surface may have unique attributes affecting cell interactions. Both the dimension and the density of the nanofeatures affect cell behavior [133]. In a well controlled investigation of titanium nanostructure, Andersson and colleagues [138] compared cell morphology and cytokine production on titanium substrates with 15 mm wide and 185 nm deep grooves vs Ti

substrates with 100 nm high, 168 nm diameter hemispherical nanopillars. The cells appeared partially aligned to the grooves and had a cytokine release similar to that found from cells on flat surfaces. Cells on hemispherical pillars had a smaller area and had more membrane projections compared to cells. Morphological changes correlated with diminished protein secretion. It has been suggested that 70–100 nm features of an implant surface are scaled to function directly with the focal adhesion of cells.

Table 3

Reported osteoblast responses to nanosurfaces – in vivo

Size/nanofeature	Tissue response	Material/fabrication	Animal/cell culture model	Ref.
3 μ m CaP coating	Elimination of tissue fibrous encapsulation and foreign body giant cell response	PLGA/CaP coated with CaP	Ratus novergicus	[127]
8 nm diameter and 100 nm length	Enhanced bone formation	PLGA mixed with Ti nanotubes	Ratus novergicus	[128]
AAT texture showed micropores and an overlapped nanometric net of filaments	Increased bone-to-implant contact	cpTi/alkali etching process with CaP solution (biomimetic)	Sheep	[126]
Not shown	Increased bone-to-implant contact	cpTi/HA – ion beam assisted deposition (IBAD)	Rabbit	[129]
~100 nm features on Ti	Increased bone-to-implant contact	cpTi/TiO ₂ blasting/HF treatment	Dog	[79]
~100 nm features on Ti	Increased differentiation	cpTi/TiO ₂ blasting/HF treatment	Ratus novergicus	[78]
Not shown	Increased osseointegration	cpTi/HA – ion beam assisted deposition (IBAD)	Dog	[90]
Discrete deposition of HA nanoparticles (20–40 nm) on Ti substrate	Increased push-out test resistance	cpTi/dual acid etch/coated with CaP by DCD	Ratus novergicus	[87]
Not shown	Increased removal torque – bone-to-implant contact – bone volume	cpTi/Sandblast/HA – ion beam assisted deposition (IBAD)	Rabbit	[96]
20–100 nm range of the features (HA)	Increased tensile test resistance	cpTi and Ti ₆ Al ₄ V/acid etch/coated with CaP by DCD	Ratus novergicus	[88]




Fig. 4. Nanoscale topography-cell interactions. There is apparent affinity of cells for nanoscale features. Here, 20–40 nm features produced by $\text{H}_2\text{O}_2/\text{H}_2\text{SO}_4$ treatment are interactive points for lamellipodia of spreading cells. The cause and effect relationship is a current point of investigation. (A) 10,000 \times image of adherent cell, (B) and (C) represent 100,000 \times images of the same adherent cell and (D) 200,000 \times magnification of the cell with nanofeatures. (B) higher magnification of the rectangle in (A). (D) higher magnification of the rectangle in (B).

Cells respond differently to the scale of roughness. Osteoprogenitor cell adhesion was enhanced on poly-L-lactide (PLLA) and polystyrene (PS) surface with nanoscale and micron-scale roughness compared to smooth surfaces. OCT-1 osteoblast-like cells grew along the surface with two different nanoscale surfaces (PLLA) and grew inside micron-scale pits of PS [139]. Similar conclusions were made when comparing nano- and micron-scale grain boundary effects on osteoblast cell adhesion and proliferation [54]. Some greater details of the relationship between surface nanofeatures and cell adhesion are emerging. Teixeira and colleagues [140] demonstrated that when cells bridge submicron-scale patterns, integrin binding was limited to substrate-adsorbed proteins on the top of the ridges. Geometrical constraints imposed by topographic features smaller than focal adhesion architecture (approximately 300 nm) actually confine the cell attachment apparatus to the top of the topographic feature. Therefore, on the nanoscale patterns, integrin occupancy within a focal adhesion may be spatially segregated whereas on microscale ridges there are no constraints on integrin–ligand binding. While the current understanding of nanotopography effects on adherent osteoblast behavior requires further clarification, nanotopography may work at a linear scale that facilitates the mechanotransduction signaling mechanisms of the adherent osteoblast.

Several investigations demonstrate that cell spreading is restricted on nanoscale surfaces. For example, Dalby and co-workers [116] investigated primary human osteoblast cell behavior on nanopitted surfaces. High pit density reduced cell spreading and ordered arrays of nanopits were effective in this regard. Randomization of the pits led to more cell spreading.

Nanotopography presents an opportunity to modulate cell adhesion and spreading both positively and negatively. When Lim and co-workers [93] compared osteoblast adhesion on PLLA substrates with 3–45 nm nanofeatures they demonstrated that cell adhesion was positively affected by nanotopography and interdependent on substratum surface characteristics of topography and surface chemistry. Lim and colleagues [94] further demonstrated that 14–29 nm pits favorably supported adherent cell integrin signaling when compared to 45 nm pits. In contrast, Cai and co-workers [141] found no major differences in fibronectin adsorption or cell proliferation on 2 vs 20 nm titanium films. There may be cell-type specific responses to nanofeatures of a given surface.

Teixeira and colleagues [142] have also shown that, depending on cell culture conditions, corneal cell integrins aligned either parallel to or perpendicular with the isotropic nanofeatures. Cellular responses to nanoscale and submicron-topographic cues are context dependent. Given the relatively anisotropic nature of natural cellular substrates, the significance of such findings remains to be defined. Nonetheless, these and other studies show that cell adhesion through integrins is sensitive to nanoscale features.

Cells adherent to nanotopographies may possess altered motility. Recent reports demonstrated that fibroblast and MSCs motility varied remarkably across a small range of nanostructures [143,144]. Hansen and colleagues [92] cultured MC3T3-E1 osteoblastic cells on nanotopographic surfaces (11–38 nm high islands). Using atomic force microscopy (AFM), they measured relatively higher cellular modulus values for cells on surfaces with nanofeatures compared with cells on flat control surfaces. They concluded that nanoscale topography affects the actual mechanical

properties of the individual cell. This may be attributed to the resultant integrin-based remodeling of the cytoskeleton or more complex biophysical changes in the cell membrane. The ability to control cell motility or spreading may be valuable in future engineering of the implant–bone–mucosa interface or the mucosa–epithelial interface at the dental implant abutment.

5.3. Proliferation

Apparently, nanoscale features can increase adherent cell proliferation. Zhao and co-workers [145] used three different approaches (electrochemical machining, anodization and chemical etching) to produce reproducible submicron-scale structures on Ti surfaces and observed an inverse relationship between cell proliferation and cell differentiation with the diminishing scale of surface features. Webster and colleagues [52] also observed increased osteoblast proliferation on the nanoscale (alumina, titania and hydroxyapatite) materials tested.

It is not fully understood how nanostructured surfaces modulate the adherent osteoblast response. At the simplest of levels, the proliferation rate of adherent cells has been measured as an index of cytocompatibility. Suggested is the concept that surface-to-cell signaling results in increased rate of proliferation. The mechanism(s) affecting this process is not defined, however, it can be speculated that many of the events associated with adhesion can affect signaling pathways that control proliferation. One example is the cross-talk between integrin signaling and the predominant MAP kinase pathways affecting cell proliferation [146].

5.4. Selectivity of adhesion

An interesting feature of nanoscale topographic surfaces is the selectivity of cell adhesion. Several investigators have demonstrated the relative diminution of fibroblast adhesion compared to osteoblast adhesion when nano- and micron-structured surfaces were evaluated [49,123]. For example, on nano-sized materials, the affinity ratio between osteoblasts and fibroblasts was 3 to 1. In the conventional materials the ratio was 1 to 1 [109]. Similar results with other cell types such as smooth muscle cells and chondrocytes have been reported [122]. This could have important implications in specification of tissue responses at bone and mucosal surfaces of the dental implant/abutment assembly.

Bacterial adhesion and proliferation is also diminished on nanophase materials [147]. Decreased bacterial colonization on nanostructured TiO₂ and ZnO is observed even though these surfaces promote osteoblast adhesion and differentiation. These initial observations imply that further development of the implant and the implant abutment surface can be explored in terms of biofilm accumulation and peri-implantitis.

The function of other cells types on nanostructured surfaces has also been addressed by Webster and co-workers [53]. They measured on nanoscale surface an increase in osteoclast function measured by tartrate resistant acid phosphatase (TRAP) synthesis and formation of resorption pits. The TRAP synthesis on nanophase hydroxyapatite was more than twice that measured on conventional hydroxyapatite. This increased osteoclastic activity may be important for the formation and maintenance of healthy new bone juxtaposed to a dental implant.

5.5. Differentiation

In addition to supporting osteoblast-specific adhesion and adherent cellular proliferation, it is important to the process of osseointegration that the adherent mesenchymal cells differentiate rapidly along the osteoblast lineage. Early indications of nanoscale topography advantages were reported by Webster et al. [51]. They

revealed that alkaline phosphatase synthesis and calcium mineral content increased in cell layers formed on nano-sized materials after 21 and 28 days.

To date few studies have evaluated the gene expression pattern indicative of differentiation of osteoblasts on nanostructured surfaces. Immunolabeled osteopontin and BSP were found in higher concentration in nanostructured surfaces [117]. Isa and co-workers [41] compared adherent palatal mesenchyme cell differentiation when cultured on a hydrophilic micron-scale topography cpTi surface or a nanoscale cpTi surface. Both surfaces supported osteoblastic differentiation, however, Runx2 expression (the key transcription factor controlling osteoblast differentiation) was increased on the nanoscale surface only. A recent in vitro and in vivo study has also demonstrated the upregulation in Runx2 expression [78]. Also, many other genes are upregulated in nanostructured surfaces as a response to Runx2 levels, such as BSP, OPN, OCN (Fig. 5).

Increased bone formation was measured for nanoscale rough implant surfaces in animal models [148]. In a series of studies the same group found early bone formation and increased torque removal when implant surfaces were added with nano-hydroxyapatite or titania [149].

6. Nanotechnology alters surface reactivity

Nanoscale modification of the implant surface may alter the endosseous implants surface reactivity. Existing reports suggest that little bone bonding occurs at endosseous titanium implants, particularly during the early phases of bone formation [150]. Nanoscale modifications of topography appear to change the chemical reactivity of bulk materials [151]. Ellingsen [152] demonstrated that the calcium phosphate precipitation on grit blasted titanium was dramatically altered by HF surface treatment that creates nanoscale topographic surface features. When the physical interaction of such titanium disks with bone was measured by a pull-off test, bonding of bone to the HF treated titanium surface was evident [153]. Bone bonding may be a benefit attributed to titanium implants through nanoscale surface modification.




Fig. 5. Effect on surface treatment (topography) on osteoblast differentiation. Osteoblasts were cultured on titanium disks treated by machining ($R_a = 86.52$ nm), acid etching to provide a micron-rough surface ($R_a = 388.40$), and with zirconia sol-gel deposition ($R_a = 89.71$ nm) to produce a nanoscale topography with pore sizes ranging from 20 to 40 nm. During the 21 days, expression of the key osteoblast differentiation factor, Runx2, was determined by real-time PCR. The results are plotted as fold change in expression level (compared to day 3 machined surface) vs culture duration (days). The marked elevation in Runx2 levels for the nanoscale surface reflects data for other nanoscale surfaces [78]. M – machined surface, Ac – acid-etched surface, and Zr – nanozirconia surface.

Biomimetic features of nanoscale modifications to the endosseous surface tissue–implant interface also address molecular (not cellular) interactions with tissues. Davies [29,150] described the formation of bone/implant bonding at solid surfaces as a four-stage process comprising the adsorption of non-collagenous bone proteins to the solid surface. Critical to the process is the initiation of mineralization by the adsorbed proteins and incipient surface directed mineralization. In a recent study, Mendes and colleagues [88] concluded that the traditional “bioactive” lithomorphous materials such as CAPs and bioactive glasses are not obligatory to promote bone bonding, but rather that a surface should have a submicron-surface complexity into which the bony cement line matrix can be deposited, and with which it can interact. Nanoscale topography may provide biomimetic surfaces that support hydroxyapatite mineral formation [154], and related organic phase guidance of bone mineralization [155].

7. The relative value of nanoscale and micron-scale roughness

The development of an implant/bone interface may be influenced by both nanoscale and micron-scale parameters of topography. The role of surface parameters (both bulk chemistry and topography) requires consideration of molecular (ionic and biomolecular) interactions with the surface, cell adhesion phenomenon and local biomechanical features of the established interface. It is clear that nanoscale modification will affect the chemical reactivity of an endosseous implant surface and alter the ionic and biomolecular interactions with the surface. Proposed changes include enhanced wettability, altered protein adsorption, and potential mineralization phenomenon. Changes in wettability and altered protein adsorption lead to altered cell adhesion, likely involving both integrin and non-integrin receptors. The potential for mineralization and epitaxial crystal growth in support of early bone bonding could dramatically alter the biomechanical environment of the healing implant in favor of stability.

Various reports support the concept that nanotopography enhances osteoblastic differentiation which could also promote stability and favorably alter the biomechanical environment for healing (see Tables 2 and 3). However, initial clinical stability may require additional considerations of micron-scale topography and overall implant design. The pioneering investigations of Meirelles and co-workers [148,149] suggest that nanometer-scale topography alone is not sufficient to assure robust osseointegration. Investigations which have isolated nanometer-scale topography as an experimental variable in osseointegration have required additional consideration of endosseous implant stability. It is possible that micron-level roughness is of additional value to the process of osseointegration. The theoretical consideration of how forming tissues interlock with micron-level topographic elements [29], and how mechanical stimulation of forming tissues is imparted by such topographic elements [28] represent ideas that may not be fully displaced by the introduction of nanotopographic modification to the endosseous implant surface.

8. Nanostructured surfaces for implant dentistry

There are many different methods to impart nanoscale features to the implant surface (see Table 1). Several of these methods have already been used to modify implants available commercially. Others are advancing through the research and development process.

As indicated above, positive bone responses occur at nanostructured surfaces tested *in vitro* and *in vivo*. Presently, only a few nanoscale surface topography modifications have been used to enhance bone responses at clinical dental implants. The OsseoSpeed surface (Astra Tech AB, Mölndal, Sweden) possesses




Fig. 6. Scanning electron microscopic evaluation of an OsseoSpeed implant. 2500× magnification of the TiO₂ grit blasted and HF treated implant surface. Note that the TiO₂ grit blasted surface is randomly covered with surface features of approximately 100 nm imparted by the HF etching.

nanostructured features created by TiO₂ blasting followed by a proprietary hydrofluoric acid treatment [44,77]. Across a micron-rough titanium surface, 50–100 nm surface accretions of titanium oxide are observed by scanning electron microscope (SEM) analysis (Fig. 6). Greater osteoblastic gene expression (Runx2, Osterix, Alkaline Phosphatase and Bone Sialoprotein) was measured in cells adherent to the nanoscale HF treated surface compared to the micron-scale surface [78]. This nanotopography is associated with the elevated levels of gene expression that indicate rapid osteoblastic differentiation. Most recent investigations show that this nanoscale surface modification promotes high levels of IGF-2 and BMP2 and BMP6 expression by adherent human mesenchymal stem cells for prolonged periods of time in culture.

Other studies concerning this nanoscale surface modification have demonstrated an increased bone formation, torque removal value [46]. In the rabbit tibia model of osseointegration, histomorphometric evaluations demonstrated higher bone-to-implant contact for the nanoscale OsseoSpeed implants compared to the micron-scale TiOblast implants (Astra Tech AB, Mölndal, Sweden) at 1 month ($35 \pm 14\%$ vs $26 \pm 8\%$) and 3 months ($39 \pm 11\%$ vs $31 \pm 6\%$) after placement. Berglundh and colleagues [79] used a gap model of osseointegration in the canine mandible to demonstrate the amount of new bone that formed in the voids within the first 2 weeks of healing was greater for HF-modified (OsseoSpeed) implants than at TiOblast implants and concluded that the nanoscale surface promotes osseointegration in the early phase of healing following implant installation.

Clinical evaluation of this implant surface preceded clinical launch and a report of the first data was provided in 2006 [18]. Six-hundred and thirty four patients received 1860 OsseoSpeed™ implants. The initial report indicated 4% surfaces had signs of inflammation (BOP) with plaque present on 12% of sites. Twenty-one patients have lost a total of 25 implants (15 in maxilla and 10 in mandible) for a CISR of 98.7% from placement. Evaluation of this effectiveness trial performed in more than 100 practices is ongoing. High success in challenging situations such as immediate placement and loading was also reported [156].

Another nanoscale surface implant presently available in the clinical marketplace involves a CaP nanoparticle modification of a minimally rough titanium alloy implant (Nanotite, 3i Implant Innovations, Palm Beach Gardens, FL). The surface has been

described as being created by a particulate sol–gel deposition method using discrete crystalline deposition (DCD) of calcium phosphate (CaP) (nominal crystal size 20 nm) with surface coverage of approximately 50%. The suggested nanofeature size of the tightly adherent adsorbed CaP/DCD crystal is 50–100 nm. Mendes and co-workers [88] measured bone ingrowth for implants modified using this technology in a rat tibia model using a well defined bone chamber model. The extent of bone ingrowth was 26.95% and 29.73% for cpTi and Ti alloy modified surfaces compared to the 12.01% cpTi and 16.97% Ti alloy chambers. In a related presentation, Mendes and colleagues [88] showed bone-bonding behavior; DCD, surfaces had statistically greater tensile detachment force (e.g. 11.30 N nanoscale DCD vs 1.90 N control).

The nanoscale CaP surface created by DCD (Nanotite, 3i) was further evaluated [157]. The histologic evaluation of clinical implants revealed bone-to-implant contact of $19 \pm 14.2\%$ and $32.2 \pm 18.5\%$ for the Osseotite (3i) control and the Nanotite (3i) experimental implants, respectively. Other clinical studies are ongoing to determine the safety and performance of this implant with nanoscale topography. For example, Goené and co-workers [158] observed greater bone formation at 4 and 8 weeks and concluded that the addition of a nanometer-scale calcium phosphate treatment to a dual acid-etched implant surface appeared to increase the extent of bone development after 4 and 8 weeks of healing. The authors suggest that this rapid accrual of bone at the implant expedites the implant healing period and supports early loading protocols.

Ion beam assisted deposition (IBAD) has also been used to create a commercially available dental implant surface [90]. This technique creates a thin-film over the implant surface by deposition of the chemical element of interest. In one available study, the bone formation (measured by tetracycline labeling quantification) was higher in the experimental group than in the control group (sand-blasted/acid-etched) after 2 (13.56% vs 24.04%) and 4 weeks (14.22% vs 27.39%) [90]. An example of this type of surface modification is presented on the Nanotite surface of Bicon Implants (Nanotite, Bicon Inc., Boston, MA). These very different chemical and physical approaches all impart nanoscale features to existing endosseous cpTitanium implant surfaces.

These initial reports of nanoscale topography implants provide insight into potential advantages for dental implant therapy. High implant survival rates have been reported. The high survival in effectiveness trials involving the HF-modified TiO₂ grit blasted surface implant and in challenging clinical examinations may reflect greater control of initial bone formation due to the rapid differentiation of osteoblastic cells observed in laboratory studies. The potential impact of bone bonding measured in preclinical studies requires further study; however, the possible advantages of bone-bonding behavior at a titanium surface could have clinical merit. How nanoscale topography and nanotechnology may be used to enhance the tissue–abutment interface remains largely unexplored. It should be noted that the currently available implants differ in their micron-level topography, in their design and in their bulk material composition. It may be difficult to derive specific conclusions from the aggregate data regarding nanoscale surface topography alone. However, for each example of current nanoscale implant surfaces of available implants, cell culture, histological, and clinical data suggest that nanoscale surfaces offer incremental advantages to clinical problems where rapid bone accrual at the implant surface provides solutions.

9. Conclusions

Nanoscale modification can alter the chemistry and/or topography of the implant surface. Different methods have been described to modify or to embellish titanium substrates with

nanoscale features. Such changes alter the implant surface interaction with ions, biomolecules and cells. These interactions can favorably influence molecular and cellular activities and alter the process of osseointegration. Cell culture studies reveal that there exists a range of nanoscale topography that promotes the osteoinductive molecular program for adherent osteoprogenitor cells. Additionally, nanoscale alterations may promote bone-bonding behavior at the titanium–bone interface. Nanoscale modification of titanium endosseous implant surfaces enhances interfacial bone formation measured as bone-to-implant contact. At this moment, both a hydrofluoric acid modified titanium endosseous implant with nanoscale features and two calcium phosphate nanofeature-modified titanium implants are available for clinical use. The potential risks and benefits of manipulating biomaterial interfaces at the nanoscale will be defined by long-term clinical evaluation of such endosseous devices.

Acknowledgements

The authors would like to thank CAPES (Coordenação de Aperfeiçoamento de Pessoal de Nível Superior) and CNPq (Conselho Nacional de Desenvolvimento Científico e Tecnológico), and UNC-Astra Tech contributions to this article by fellowship.

References

- [1] Branemark PI, Adell R, Breine U, Hansson BO, Lindström J, Ohlsson A. Intraosseous anchorage of dental prostheses. I. Experimental studies. *Scand J Plast Reconstr Surg* 1969;3:81–100.
- [2] Linder L, Albrektsson T, Brånemark PI, Hansson HA, Ivarsson B, Jönsson U, et al. Electron microscopic analysis of the bone–titanium interface. *Acta Orthop Scand* 1983;54:45–52.
- [3] Albrektsson T, Sennarby L. Direct bone anchorage of oral implants: clinical and experimental considerations of the concept of osseointegration. *Int J Prosthodont* 1990;3:30–41.
- [4] Adell R, Eriksson B, Lekholm U, Branemark PI, Jemt T. Long-term follow-up study of osseointegrated implants in the treatment of totally edentulous jaws. *Int J Oral Maxillofac Implants* 1990;5:347–59.
- [5] Adell R, Lekholm U, Rockler B, Branemark PI. A 15-year study of osseointegrated implants in the treatment of the edentulous jaw. *Int J Oral Surg* 1981;10:387–416.
- [6] Albrektsson T, Dahl E, Enbom L, Engvall S, Engquist B, Eriksson AR, et al. Osseointegrated oral implants. A Swedish multicenter study of 8139 consecutively inserted Nobelpharma implants. *J Periodontol* 1988;59:287–96.
- [7] Goodacre CJ, Kan JY, Rungcharassaeng K. Clinical complications of osseointegrated implants. *J Prosthet Dent* 1999;81:537–52.
- [8] Zarb GA, Schmitt A. The longitudinal clinical effectiveness of osseointegrated dental implants: the Toronto study. Part III: problems and complications encountered. *J Prosthet Dent* 1990;64:185–94.
- [9] Morton D, Jaffin R, Weber HP. Immediate restoration and loading of dental implants: clinical considerations and protocols. *Int J Oral Maxillofac Implants* 2004;19(Suppl.):103–8.
- [10] Tolstunov L. Dental implant success–failure analysis: a concept of implant vulnerability. *Implant Dent* 2006;15:341–6.
- [11] Jaffin RA, Berman CL. The excessive loss of Branemark fixtures in type IV bone: a 5-year analysis. *J Periodontol* 1991;62:2–4.
- [12] Bain CA. Smoking and implant failure–benefits of a smoking cessation protocol. *Int J Oral Maxillofac Implants* 1996;11:756–9.
- [13] Fiorellini JP, Chen PK, Nevins M, Nevins ML. A retrospective study of dental implants in diabetic patients. *Int J Periodontics Restorative Dent* 2000;20:366–73.
- [14] Cooper LF. Biologic determinants of bone formation for osseointegration: clues for future clinical improvements. *J Prosthet Dent* 1998;80:439–49.
- [15] Nanci A, Wuest JD, Peru L, Brunet P, Sharma V, Zalzal S, et al. Chemical modification of titanium surfaces for covalent attachment of biological molecules. *J Biomed Mater Res* 1998;40:324–35.
- [16] Boyan BD, Schwartz Z, Hambleton JC. Response of bone and cartilage cells to biomaterials in vivo and in vitro. *J Oral Implantol* 1993;19:116–22.
- [17] Schwartz Z, Swain LD, Marshall T, Sela J, Gross U, Amir D, et al. Modulation of matrix vesicle enzyme activity and phosphatidylserine content by ceramic implant materials during endosteal bone healing. *Calcif Tissue Int* 1992;51:429–37.
- [18] Stanford CM, Johnson GK, Fakhry A, Gratton D, Mellonig JT, Wanger W. Outcomes of a fluoride modified implant one year after loading in the posterior-maxilla when placed with the osteotome surgical technique. *Appl Osseointegration Res* 2006;5:50–5.

- [19] Masuda T, Salvi GE, Offenbacher S, Felton DA, Cooper LF. Cell and matrix reactions at titanium implants in surgically prepared rat tibiae. *Int J Oral Maxillofac Implants* 1997;12:472–85.
- [20] Meyer U, Joos U, Mythili J, Stamm T, Hohoff A, Fillies T, et al. Ultrastructural characterization of the implant/bone interface of immediately loaded dental implants. *Biomaterials* 2004;25:1959–67.
- [21] Berglundh T, Abrahamsson I, Lang NP, Lindhe J. De novo alveolar bone formation adjacent to endosseous implants. *Clin Oral Implants Res* 2003;14: 251–62.
- [22] Le Guéhennec L, Soueidan A, Layrolle P, Amouriq Y. Surface treatments of titanium dental implants for rapid osseointegration. *Dent Mater* 2007;23: 844–54.
- [23] Kasemo B. Biocompatibility of titanium implants: surface science aspects. *J Prosthet Dent* 1983;49:832–7.
- [24] Johansson CB, Albrektsson T. A removal torque and histomorphometric study of commercially pure niobium and titanium implants in rabbit bone. *Clin Oral Implants Res* 1991;2:24–9.
- [25] Suska F, Gretzer C, Esposito M, Emanuelsson L, Wennerberg A, Tengvall P, et al. In vivo cytokine secretion and NF-kappaB activation around titanium and copper implants. *Biomaterials* 2005;26:519–27.
- [26] Buser D, Schenk RK, Steinemann S, Fiorellini JP, Fox CH, Stich H. Influence of surface characteristics on bone integration of titanium implants. A histomorphometric study in miniature pigs. *J Biomed Mater Res* 1991;25: 889–902.
- [27] Gotfredsen K, Hjorting-Hansen E, Budtz-Jørgensen E. Clinical and radiographic evaluation of submerged and nonsubmerged implants in monkeys. *Int J Prosthodont* 1990;3:463–9.
- [28] Hansson S, Norton M. The relation between surface roughness and interfacial shear strength for bone-anchored implants. A mathematical model. *J Biomech* 1999;32:829–36.
- [29] Davies JE. Understanding peri-implant endosseous healing. *J Dent Educ* 2003;67:932–49.
- [30] Puleo DA, Nanci A. Understanding and controlling the bone-implant interface. *Biomaterials* 1999;20:2311–21.
- [31] Schwartz Z, Lohmann CH, Oefinger J, Bonewald LF, Dean DD, Boyan BD. Implant surface characteristics modulate differentiation behavior of cells in the osteoblastic lineage. *Adv Dent Res* 1999;13:38–48.
- [32] Albrektsson T, Wennerberg A. Oral implant surfaces: part 1 – review focusing on topographic and chemical properties of different surfaces and in vivo responses to them. *Int J Prosthodont* 2004;17:536–43.
- [33] Albrektsson T, Wennerberg A. Oral implant surfaces: part 2 – review focusing on clinical knowledge of different surfaces. *Int J Prosthodont* 2004;17:544–64.
- [34] Burger EH, Klein-Nulend J. Mechanotransduction in bone – role of the lacuno-canalicular network. *FASEB J* 1999;13(Suppl.):S101–12.
- [35] Hansson S. The dental implant meets bone – a clash of two paradigms. *Appl Osseointegration Res* 2006;1:15–7.
- [36] Wong M, Eulenberger J, Schenk R, Hunziker E. Effect of surface topology on the osseointegration of implant materials in trabecular bone. *J Biomed Mater Res* 1995;29:1567–75.
- [37] Wennerberg A, Ektesabi A, Albrektsson T, Johansson C, Andersson B. A 1-year follow-up of implants of differing surface roughness placed in rabbit bone. *Int J Oral Maxillofac Implants* 1997;12:486–94.
- [38] Park JY, Gemmell CH, Davies JE. Platelet interactions with titanium: modulation of platelet activity by surface topography. *Biomaterials* 2001; 22:J671–82.
- [39] Ricci JL, Grew JC, Alexander H. Connective-tissue responses to defined biomaterial surfaces. I. Growth of rat fibroblast and bone marrow cell colonies on microgrooved substrates. *J Biomed Mater Res A* 2007 Aug 9 [Epub ahead of print].
- [40] Schneider GB, Perinpanayagam H, Clegg M, Zaharias R, Seabold D, Keller J, et al. Implant surface roughness affects osteoblast gene expression. *J Dent Res* 2003;82:372–6.
- [41] Isa ZM, Schneider GB, Zaharias R, Seabold D, Stanford CM. Effects of fluoride modified titanium surfaces on osteoblast proliferation and gene expression. *Int J Oral Maxillofac Implants* 2006;21:203–11.
- [42] Ogawa T, Nishimura I. Different bone integration profiles of turned and acid-etched implants associated with modulated expression of extracellular matrix genes. *Int J Oral Maxillofac Implants* 2003;18:200–10.
- [43] Ogawa T, Nishimura I. Genes differentially expressed in titanium implant healing. *J Dent Res* 2006;85:566–70.
- [44] Abron A, Hopfensperger M, Thompson J, Cooper LF. Evaluation of a predictive model for implant surface topography effects on early osseointegration in the rat tibia model. *J Prosthet Dent* 2001;85:40–6.
- [45] Buser D, Broggini N, Wieland M, Schenk RK, Denzer AJ, Cochran DL, et al. Enhanced bone apposition to a chemically modified SLA titanium surface. *J Dent Res* 2004;83:529–33.
- [46] Ellingsen JE, Johansson CB, Wennerberg A, Holmen A. Improved retention and bone-to-implant contact with fluoride-modified titanium implants. *Int J Oral Maxillofac Implants* 2004;19:659–66.
- [47] Gutwein LG, Webster TJ. Increased viable osteoblast density in the presence of nanophase compared to conventional alumina and titania particles. *Biomaterials* 2004;25:4175–83.
- [48] Oh SH, Finones RR, Daraio C, Chen LH, Jin S. Growth of nano-scale hydroxyapatite using chemically treated titanium oxide nanotubes. *Biomaterials* 2005;26:4938–43.
- [49] Price RL, Gutwein LG, Kaledin L, Tepper F, Webster TJ. Osteoblast function on nanophase alumina materials: influence of chemistry, phase, and topography. *J Biomed Mater Res A* 2003;67:1284–93.
- [50] Price RL, Haberstrohm KM, Webster TJ. Enhanced functions of osteoblasts on nanostructured surfaces of carbon and alumina. *Med Biol Eng Comput* 2003; 41:372–5.
- [51] Webster TJ, Siegel RW, Bizios R. Osteoblast adhesion on nanophase ceramics. *Biomaterials* 1999;20:1221–7.
- [52] Webster TJ, Ergun C, Doremus RH, Siegel RW, Bizios R. Enhanced functions of osteoblasts on nanophase ceramics. *Biomaterials* 2000;21:1803–10.
- [53] Webster TJ, Schadler LS, Siegel RW, Bizios R. Mechanisms of enhanced osteoblast adhesion on nanophase alumina involve vitronectin. *Tissue Eng* 2001;7:291–301.
- [54] Webster TJ, Eijffor JU. Increased osteoblast adhesion on nanophase metals: Ti, Ti₆Al₄V, and CoCrMo. *Biomaterials* 2004;25:4731–9.
- [55] Schwartz Z, Nasazky E, Boyan BD. Surface microtopography regulates osteointegration: the role of implant surface microtopography in osteointegration. *Alpha Omega* 2005;98:9–19.
- [56] Webster TJ, Hellenmeyer EL, Price RL. Increased osteoblast functions on theta + delta nanofiber alumina. *Biomaterials* 2005;26:953–60.
- [57] Zhao G, Schwartz Z, Wieland M, Rupp F, Geis-Gerstorf J, Cochran DL, et al. High surface energy enhances cell response to titanium substrate microstructure. *J Biomed Mater Res A* 2005;74:49–58.
- [58] Dike LE, Chen CS, Mrksich M, Tien J, Whitesides GM, Ingber DE. Geometric control of switching between growth, apoptosis, and differentiation during angiogenesis using micropatterned substrates. *In Vitro Cell Dev Biol Anim* 1999;35:441–8.
- [59] García AJ, Reyes CD. Bio-adhesive surfaces to promote osteoblast differentiation and bone formation. *J Dent Res* 2005;84:407–13.
- [60] Trisi P, Lazzara R, Rebaudi A, Rao W, Testori T, Porter SS. Bone-implant contact on machined and dual acid-etched surfaces after 2 months of healing in the human maxilla. *J Periodontol* 2003;74:945–56.
- [61] Wang L, Zhao G, Olivares-Navarrete R, Bell BF, Wieland M, Cochran DL, et al. Integrin beta1 silencing in osteoblasts alters substrate-dependent responses to 1,25-dihydroxy vitamin D3. *Biomaterials* 2006;27:3716–25.
- [62] Cochran DL. A comparison of endosseous dental implant surfaces. *J Periodontol* 1999;70:1523–39.
- [63] Shalabi MM, Gortemaker A, Van't Hof MA, Jansen JA, Creugers NH. Implant surface roughness and bone healing: a systematic review. *J Dent Res* 2006; 85:496–500.
- [64] Becker J, Kirsch A, Schwarz F, Chatzinikolaïdou M, Rothamel D, Lekovic V, et al. Bone apposition to titanium implants bio-coated with recombinant human bone morphogenetic protein-2 (rhBMP-2). A pilot study in dogs. *Clin Oral Investig* 2006;10:217–24.
- [65] Schliephake H, Aref A, Scharnweber D, Bierbaum S, Roessler S, Sewing A. Effect of immobilized bone morphogenic protein 2 coating of titanium implants on peri-implant bone formation. *Clin Oral Implants Res* 2005;16:563–9.
- [66] Christensen EM, Anseth KS, van den Beucken JJ, Chan CK, Ercan B, Jansen JA, et al. Nanobiomaterial applications in orthopedics. *J Orthop Res* 2007;25:11–22.
- [67] Liu H, Slamovich EB, Webster TJ. Increased osteoblast functions among nanophase titania/poly(lactide-co-glycolide) composites of the highest nanometer surface roughness. *J Biomed Mater Res A* 2006;78:798–807.
- [68] Scotchford CA, Gilmore CP, Cooper E, Leggett GJ, Downes S. Protein adsorption and human osteoblast-like cell attachment and growth on alkylthiol on gold self-assembled monolayers. *J Biomed Mater Res* 2002;59:84–99.
- [69] Germanier Y, Tosatti S, Broggini N, Textor M, Buser D. Enhanced bone apposition around biofunctionalized sandblasted and acid-etched titanium implant surfaces. A histomorphometric study in miniature pigs. *Clin Oral Implants Res* 2006;17:251–7.
- [70] Zhou J, Chang C, Zhang R, Zhang L. Hydrogels prepared from unsubstituted cellulose in NaOH/urea aqueous solution. *Macromol Biosci* 2007;7:804–9.
- [71] Kim HM, Kokubo T, Fujibayashi S, Nishiguchi S, Nakamura T. Bioactive macroporous titanium surface layer on titanium substrate. *J Biomed Mater Res* 2000;5(52):553–7.
- [72] Wang XX, Hayakawa S, Tsuru K, Osaka A. A comparative study of in vitro apatite deposition on heat-, H₂O(2)-, and NaOH-treated titanium surfaces. *J Biomed Mater Res* 2001;54:172–8.
- [73] Uchida M, Kim HM, Miyaji F, Kokubo T, Nakamura T. Apatite formation on zirconium metal treated with aqueous NaOH. *Biomaterials* 2002;23:313–7.
- [74] Wang XX, Hayakawa S, Tsuru K, Osaka A. Bioactive titania-gel layers formed by chemical treatment of Ti substrate with a H₂O₂/HCl solution. *Biomaterials* 2002;23:1353–7.
- [75] Mante FK, Little K, Mante MO, Rawle C, Baran GR. Oxidation of titanium, RGD peptide attachment, and matrix mineralization rat bone marrow stromal cells. *J Oral Implantol* 2004;30:343–9.
- [76] Ellingsen JE, Thomsen P, Lyngstadaas SP. Advances in dental implant materials and tissue regeneration. *Periodontology* 2000;2006(41):136–56.
- [77] Cooper LF, Zhou Y, Takebe J, Guo J, Abron A, Holmen A, et al. Fluoride modification effects on osteoblast behavior and bone formation at TiO₂ grit blasted c.p. titanium endosseous implants. *Biomaterials* 2006;27: 926–36.
- [78] Guo J, Padilla RJ, Ambrose W, De Kok JJ, Cooper LF. Modification of TiO₂ grit blasted titanium implants by hydrofluoric acid treatment alters adherent osteoblast gene expression in vitro and in vivo. *Biomaterials* 2007; 28:5418–25.

- [79] Berglund T, Abrahamsson I, Albouy JP, Lindhe J. Bone healing at implants with a fluoride-modified surface: an experimental study in dogs. *Clin Oral Implants Res* 2007;18:147–52.
- [80] Ben-Nissan B, Choi AH. Sol-gel production of bioactive nanocoatings for medical applications. Part 1: an introduction. *Nanomed* 2006;1:311–9.
- [81] Liu DM, Troczynski T, Tseng WJ. Water-based sol-gel synthesis of hydroxyapatite: process development. *Biomaterials* 2001;22:1721–30.
- [82] Kim HW, Koh YH, Li LH, Lee S, Kim HE. Hydroxyapatite coating on titanium substrate with titania buffer layer processed by sol-gel method. *Biomaterials* 2004;25:2533–8.
- [83] Lee SH, Kim HW, Lee EJ, Li LH, Kim HE. Hydroxyapatite-TiO₂ hybrid coating on Ti implants. *J Biomater Appl* 2006;20:195–208.
- [84] Piveteau LD, Gasser B, Schlapbach L. Evaluating mechanical adhesion of sol-gel titanium dioxide coatings containing calcium phosphate for metal implant application. *Biomaterials* 2000;21:2193–201.
- [85] Arias JL, Mayor MB, Pou J, Leng Y, León B, Pérez-Amor M. Micro- and nano-testing of calcium phosphate coatings produced by pulsed laser deposition. *Biomaterials* 2003;24:3403–8.
- [86] Choi AH, Ben-Nissan B. Sol-gel production of bioactive nanocoatings for medical applications. Part II: current research and development. *Nanomed* 2007;2:51–61.
- [87] Nishimura I, Huang Y, Butz F, Ogawa T, Lin L, JakeWang C. Discrete deposition of hydroxyapatite nanoparticles on a titanium implant with predisposing substrate microtopography accelerated osseointegration. *Nanotechnology* 2007;18:245101 (9pp).
- [88] Mendes VC, Moineddin R, Davies JE. The effect of discrete calcium phosphate nanocrystals on bone-bonding to titanium surfaces. *Biomaterials* 2007;28:4748–55.
- [89] Ricci JL, Charvet J, Frenkel SR, Chang R, Nadkarni P, Turner J, et al. Bone response to laser microtextured surfaces. In: Davies JE, editor. *Bone engineering*. Toronto: Em2 Inc.; 2000. p. 8–9 [Chapter 25].
- [90] Coelho PG, Suzuki M. Evaluation of an ibad thin-film process as an alternative method for surface incorporation of bioceramics on dental implants. A study in dogs. *J Appl Oral Sci* 2005;13:87–92.
- [91] Brody S, Anilkumar T, Liliensiek S, Last JA, Murphy CJ, Pandit A. Characterizing nanoscale topography of the aortic heart valve basement membrane for tissue engineering heart valve scaffold design. *Tissue Eng* 2006;12:413–21.
- [92] Hansen JC, Lim JY, Xu LC, Siedlecki CA, Mauger DT, Donahue HJ. Effect of surface nanoscale topography on elastic modulus of individual osteoblastic cells as determined by atomic force microscopy. *J Biomech* 2007;40:2865–71.
- [93] Lim JY, Hansen JC, Siedlecki CA, Runt J, Donahue HJ. Human foetal osteoblastic cell response to polymer-demixed nanotopographic interfaces. *J R Soc Interface* 2005;2:97–108.
- [94] Lim JY, Dreiss AD, Zhou Z, Hansen JC, Siedlecki CA, Hengstebeck RW, et al. The regulation of integrin-mediated osteoblast focal adhesion and focal adhesion kinase expression by nanoscale topography. *Biomaterials* 2007;28:1787–97.
- [95] Palin E, Liu H, Webster TJ. Mimicking the nanofeatures of bone increases bone-forming cell adhesion and proliferation. *Nanotechnology* 2005;16:1828–35.
- [96] Park YS, Yi KY, Lee IS, Han CH, Jung YC. The effects of ion beam-assisted deposition of hydroxyapatite on the grit-blasted surface of endosseous implants in rabbit tibiae. *Int J Oral Maxillofac Implants* 2005;20:31–8.
- [97] Klabunde KJ, Strak J, Koper O, Mohs C, Park D, Decker S, et al. Nanocrystals as stoichiometric reagents with unique surface chemistry. *J Phys Chem* 1996;100:12141.
- [98] Wu SJ, DeJong LC, Rahaman MN. Sintering of nanophase γ -Al₂O₃ powder. *J Am Ceram Soc* 1996;79:2207.
- [99] Baraton MI, Chen X, Gonsalves KE. FTIR study of nanostructured alumina nitride powder surface: determination of the acidic/basic sites by CO, CO₂, and acetic acid adsorptions. *Nanostruct Mater* 1997;8:435.
- [100] Zreiqat H, Valenzuela SM, Nissán BB, Roest R, Knabe C, Radlanski RJ, et al. The effect of surface chemistry modification of titanium alloy on signalling pathways in human osteoblasts. *Biomaterials* 2005;26:7579–86.
- [101] Monsees TK, Barth K, Tippelt S, Heide K, Gorbunov A, Pompe W, et al. Effects of different titanium alloys and nanosize surface patterning on adhesion, differentiation, and orientation of osteoblast-like cells. *Cells Tissues Organs* 2005;180:81–95.
- [102] Hart A, Gadegaard N, Wilkinson CDW, Oreffo ROC, Dalby MJ. Filopodial sensing of nanotopography in osteoprogenitor cells. *Eur Cell Mater* 2005;10(Suppl. 2):65.
- [103] Hart A, Gadegaard N, Wilkinson CDW, Oreffo ROC, Dalby MJ. Osteoprogenitor response to low-adhesion nanotopographies originally fabricated by electron beam lithography. *J Mater Sci Mater Med* 2007;18:1211–8.
- [104] Washburn NR, Yamada KM, Simon Jr CG, Kennedy SB, Amis EJ. High-throughput investigation of osteoblast response to polymer crystallinity: influence of nanometer-scale roughness on proliferation. *Biomaterials* 2004;25:1215–24.
- [105] Eisenbarth E, Velten D, Breme J. Biomimetic implant coatings. *Biomol Eng* 2007;24:27–32.
- [106] Webster TJ, Ergun C, Doremus RH, Lanford WA. Increased osteoblast adhesion on titanium-coated hydroxylapatite that forms CaTiO₃. *J Biomed Mater Res A* 2003;67:975–80.
- [107] Ergun C, Liu H, Halloran JW, Webster TJ. Increased osteoblast adhesion on nanograined hydroxyapatite and tricalcium phosphate containing calcium titanate. *J Biomed Mater Res A* 2007;80:990–7.
- [108] Balasundaram G, Sato M, Webster TJ. Using hydroxyapatite nanoparticles and decreased crystallinity to promote osteoblast adhesion similar to functionalizing with RGD. *Biomaterials* 2006;27:2798–805.
- [109] Webster TJ, Ergun C, Doremus RH, Siegel RW, Bizios R. Specific proteins mediate enhanced osteoblast adhesion on nanophase ceramics. *J Biomed Mater Res* 2000;51:475–83.
- [110] Chun AL, Moralez JG, Webster TJ, Fenniri H. Helical rosette nanotubes: a biomimetic coating for orthopedics? *Biomaterials* 2005;26:7304–9.
- [111] Dong W, Zhang T, Epstein J, Cooney L, Wang H, Li Y, et al. Multifunctional nanowire bioscaffolds on titanium. *Chem Mater* 2007;19:4454–9.
- [112] Advincula MC, Rahemtulla FG, Advincula RC, Ada ET, Lemons JE, Bellis SL. Osteoblast adhesion and matrix mineralization on sol-gel-derived titanium oxide. *Biomaterials* 2006;27:2201–12.
- [113] Webster TJ, Smith TA. Increased osteoblast function on PLGA composites containing nanophase titania. *J Biomed Mater Res A* 2005;74:677–86.
- [114] Popat KC, Leoni L, Grimes CA, Desai TA. Influence of engineered titania nanotubular surfaces on bone cells. *Biomaterials* 2007;28:3188–97.
- [115] Popat KC, Chatvanichkul KI, Barnes GL, Latempa Jr TJ, Grimes CA, Desai TA. Osteogenic differentiation of marrow stromal cells cultured on nanoporous alumina surfaces. *J Biomed Mater Res A* 2007;80:955–64.
- [116] Dalby MJ, McCloy D, Robertson M, Wilkinson CD, Oreffo RO. Osteoprogenitor response to defined topographies with nanoscale depths. *Biomaterials* 2006;27:1306–15.
- [117] Oliveira PT, Nanci A. Nanotexturing of titanium-based surfaces upregulates expression of bone sialoprotein and osteopontin by cultured osteogenic cells. *Biomaterials* 2004;25:403–13.
- [118] Bigi A, Nicoli-Aldini N, Bracci B, Zavan B, Boanini E, Sbaiz F, et al. In vitro culture of mesenchymal cells onto nanocrystalline hydroxyapatite-coated Ti₁₃Nb₁₃Zr alloy. *J Biomed Mater Res A* 2007;82:213–21.
- [119] Tan J, Saltzman WM. Biomaterials with hierarchically defined micro- and nanoscale structure. *Biomaterials* 2004;25:3593–601.
- [120] Oliveira PT, Zalzal SF, Beloti MM, Rosa AL, Nanci A. Enhancement of in vitro osteogenesis on titanium by chemically produced nanotopography. *J Biomed Mater Res A* 2007;80:554–64.
- [121] Berry CC, Dalby MJ, Oreffo RO, McCloy D, Affrosman S. The interaction of human bone marrow cells with nanotopographical features in three dimensional constructs. *J Biomed Mater Res A* 2006;79:431–9.
- [122] Price RL, Ellison K, Haberstroh KM, Webster TJ. Nanometer surface roughness increases select osteoblast adhesion on carbon nanofiber compacts. *J Biomed Mater Res A* 2004;70:129–38.
- [123] McManus AJ, Doremus RH, Siegel RW, Bizios R. Evaluation of cytocompatibility and bending modulus of nanoceramic/polymer composites. *J Biomed Mater Res A* 2005;72:98–106.
- [124] Kay S, Thapa A, Haberstroh KM, Webster TJ. Nanostructured polymer/nanophase ceramic composites enhance osteoblast and chondrocyte adhesion. *Tissue Eng* 2002;8:753–61.
- [125] Briggs EP, Walpole AR, Wilshaw PR, Karlsson M, Palsgard E. Formation of highly adherent nano-porous alumina on Ti-based substrates: a novel bone implant coating. *J Mater Sci Mater Med* 2004;15:1021–9.
- [126] Chiesa R, Giavaresi G, Fini M, Sandrini E, Giordano C, Bianchi A, et al. In vitro and in vivo performance of a novel surface treatment to enhance osseointegration of endosseous implants. *Oral Surg Oral Med Oral Pathol Oral Radiol Endod* 2007;103:745–56.
- [127] Lickorish D, Guan L, Davies JE. A three-phase, fully resorbable, polyester/calcium phosphate scaffold for bone tissue engineering: evolution of scaffold design. *Biomaterials* 2007;28:1495–502.
- [128] Kubota S, Johkura K, Asanuma K, Okouchi Y, Ogiwara N, Sasaki K, et al. Titanium oxide nanotubes for bone regeneration. *J Mater Sci Mater Med* 2004;15:1031–5.
- [129] Jung YC, Han CH, Lee IS, Kim HE. Effects of ion beam-assisted deposition of hydroxyapatite on the osseointegration of endosseous implants in rabbit tibiae. *Int J Oral Maxillofac Implants* 2001;16:809–18.
- [130] Fath K, Edgell C, Burridge K. The distribution of distinct integrins in focal contacts is determined by the substratum composition. *J Cell Sci* 1989;92:67–75.
- [131] Tosatti S, Schwartz Z, Campbell C, Cochran DL, VandeVondele S, Hubbell JA, et al. RGD-containing peptide GCRGYGRGDSPG reduces enhancement of osteoblast differentiation by poly(L-lysine)-graft-poly(ethylene glycol)-coated titanium surfaces. *J Biomed Mater Res A* 2004;68:458–72.
- [132] Sinha RK, Tuan RS. Regulation of human osteoblast integrin expression by orthopedic implant materials. *Bone* 1996;18:451–7.
- [133] Cavalcanti-Adam EA, Volberg T, Micoulet A, Kessler H, Geiger B, Spatz JP. Cell spreading and focal adhesion dynamics are regulated by spacing of integrin ligands. *Biophys J* 2007;92:2964–74.
- [134] Rodrigues SN, Gonçalves IC, Martins MC, Barbosa MA, Ratner BD. Fibrinogen adsorption, platelet adhesion and activation on mixed hydroxyl-/methyl-terminated self-assembled monolayers. *Biomaterials* 2006;27:5357–67.
- [135] Arima Y, Iwata H. Effect of wettability and surface functional groups on protein adsorption and cell adhesion using well-defined mixed self-assembled monolayers. *Biomaterials* 2007;28:3074–82.
- [136] Park GE, Webster TJ. A review of nanotechnology for the development of better orthopedic implants. *J Biomed Nanotechnol* 2005;1:18–29.
- [137] Brunette DM. The effects of implant surface topography on the behavior of cells. *Int J Oral Maxillofac Implants* 1988;3:231–4.
- [138] Andersson AS, Bäckhed F, von Euler A, Richter-Dahlfors A, Sutherland D, Kasemo B. Nanoscale features influence epithelial cell morphology and cytokine production. *Biomaterials* 2003;24:3427–36.

- [139] Wan Y, Wang Y, Liu Z, Qu X, Han B, Bei J, et al. Adhesion and proliferation of OCT-1 osteoblast-like cells on micro- and nano-scale topography structured poly(L-lactide). *Biomaterials* 2005;26:4453–9.
- [140] Teixeira AI, Abrams GA, Bertics PJ, Murphy CJ, Nealey PF. Epithelial contact guidance on well-defined micro- and nanostructured substrates. *J Cell Sci* 2003;116:1881–92.
- [141] Cai K, Bossert J, Jandt KD. Does the nanometre scale topography of titanium influence protein adsorption and cell proliferation? *Colloids Surf B Bio-interfaces* 2006;49:136–44.
- [142] Teixeira AI, McKie GA, Foley JD, Bertics PJ, Nealey PF, Murphy CJ. The effect of environmental factors on the response of human corneal epithelial cells to nanoscale substrate topography. *Biomaterials* 2006;27:3945–54.
- [143] Alsberg E, Feinstein E, Joy MP, Prentiss M, Ingber DE. Magnetically-guided self-assembly of fibrin matrices with ordered nano-scale structure for tissue engineering. *Tissue Eng* 2006;12:3247–56.
- [144] Dalby MJ, Gadegaard N, Tare R, Andar A, Riehle MO, Herzyk P, et al. The control of human mesenchymal cell differentiation using nanoscale symmetry and disorder. *Nat Mater* 2007;6:997–1003.
- [145] Zhao G, Zinger O, Schwartz Z, Wieland M, Landolt D, Boyan BD. Osteoblast-like cells are sensitive to submicron-scale surface structure. *Clin Oral Implants Res* 2006;17:258–64.
- [146] Schwartz MA, Ginsberg MH. Networks and crosstalk: integrin signalling spreads. *Nat Cell Biol* 2002;4:E65–8.
- [147] Colon G, Ward BC, Webster TJ. Increased osteoblast and decreased *Staphylococcus epidermidis* functions on nanophase ZnO and TiO₂. *J Biomed Mater Res A* 2006;78:595–604.
- [148] Meirelles L, Arvidsson A, Albrektsson T, Wennerberg A. Increased bone formation to unstable nano rough titanium implants. *Clin Oral Implants Res* 2007;18:326–32.
- [149] Meirelles L. On nano size structures for enhanced early bone formation [Ph.D.]. Gothenburg: Gothenburg University; 2007.
- [150] Davies JE. Bone bonding at natural and biomaterial surfaces. *Biomaterials* 2007;28:5058–67.
- [151] Tasker LH, Sparey-Taylor GJ, Nokes LD. Applications of nanotechnology in orthopaedics. *Clin Orthop Relat Res* 2007;456:243–9.
- [152] Ellingsen JE. On the properties of surface-modified titanium. In: Davies JE, editor. *Bone engineering*. Toronto, Canada: em Squared Inc.; 2000. p. 183–8.
- [153] Ellingsen JE, Lyngstadaas SP. Increasing biocompatibility by chemical modification of titanium surfaces. In: Ellingsen JE, Lyngstadaas PS, editors. *Bio-implant interface: improving biomaterials and tissue reactions*. Boca Raton, Florida: CRC Press LLC; 2003. p. 323–40.
- [154] Ward BC, Webster TJ. The effect of nanotopography on calcium and phosphorus deposition on metallic materials in vitro. *Biomaterials* 2006;27:3064–74.
- [155] Zhu B, Lu Q, Yin J, Hu J, Wang Z. Alignment of osteoblast-like cells and cell-produced collagen matrix induced by nanogrooves. *Tissue Eng* 2005;11:825–34.
- [156] Oxby G, Lindqvist J, Nilsson P. Early loading of Astra Tech OsseoSpeed implants placed in thin alveolar ridges and fresh extraction sockets. *Appl Osseointegration Res* 2006;5:68–72.
- [157] Orsini G, Piattelli M, Scarano A, Petrone G, Kenealy J, Piattelli A, et al. Randomized, controlled histologic and histomorphometric evaluation of implants with nanometer-scale calcium phosphate added to the dual acid-etched surface in the human posterior maxilla. *J Periodontol* 2007;78:209–18.
- [158] Goené RJ, Testori T, Trisi P. Influence of a nanometer-scale surface enhancement on de novo bone formation on titanium implants: a histomorphometric study in human maxillae. *Int J Periodontics Restorative Dent* 2007;27:211–9.